

## **Technical Data Sheet** General Purpose Refrigerator 700L

Revision-A

	Model Number				
Specifications	TSG700REGBV				
	Application, Rating and I				
Application	Storage of General (non-flammable) Laboratory Materials				
Storage Volume	620Liters / 21.9 Cu	bic Feet			
Factory Set Point	+5 °C				
Set Point / Increment	+5 °C / 0.1 °C				
Electrical Power / Rated Current	230V 50 Hz 1 Phase / 1.6A				
Building Supply Rating	16A 250V	(40.5)			
Power Plug/Power Cord Length	16 Amp 250V / 4 Mete	ers (13 feet)			
Agency Listings	CE				
Standards and directives	2006/42/EC, 2014/30/EU, 2014/35/EU, 2011/6				
Electromagnetic Compatibility	Emissions (Class A), Immunity (II				
Indoor/Outdoor Usage	Indoor Use O				
Application Environment	Non-Corrosive, Non-Flammable, Non-Ex				
Ambient Operating Temperature	15° C to 32° C (59° F				
	Refrigeration Confi				
Refrigeration System	Vapor compression				
Compressor / Number / Capacity	Single speed / 1 /				
Condenser Type/Number	Forced-Air Cooled Finned-Tube C				
Expansion Device	Capillary Tub				
Evaporator Type / Defrost Method	Forced Air Finned-Tube Coil Heat Exc				
Refrigerant Charge / Flammability	R290,90gms,GWP:3 ODP				
	Controller/Electrical System Configuration and Features				
Controller Level	Eye Level				
Power Switch	On the Controller Main Screen				
Controller Type	Electronic Controller with LCD display				
Setpoint Security / Programmable	Yes / Yes				
Compressor Safe Guard	Overload Protectio				
Control Sensor	NTC Probe				
Connectivity / Remote OutputsTerminals	Dry Contact (NO a	·			
Adjustable Warm / Cold Alarms	Fully Adjustat				
Power Failure Alarm	Acoustic and Vi				
Electronic Chart Recorder	Not Provide	=-			
	Dimensions and Cor				
Interior Dimensions (H x D x W)	1505 mm x 686 mm x 600 mm (				
Exterior Dimensions (H x D x W)	2085 mm x 840 mm x 750 mm (				
Insulation	Polyurethane Foam 75	5 MM (2.95")			
Door Perimeter heater	No				
Shelves / Load Limit	4 Shelves / 30kg (66 lbs) per Shelf				
Casters	2 Lockable Castors at the front & 2	Fixed Castors at the rear			
Non-Crated Shipping Weight	162kgs / 357.1 lbs				
	Typical Performance Ch	naracteristics			
Set Point 5°C @ 20°C Ambient	Set Point 5°C @ 20°C Ambient	Test Unit Series Number or MSO Number: 20297-A-8			
20 — Pulldown — Warmup	16MINAVG	Cabinet Load: Unloader			
20 — Pulldown — Warmup 18 17 17 16 15 14 11 11 19 19 19 19 19 19 19 19 19 19 19		Average Cabinet Temp (C): 5.22			
17 16 16 17 18 18 18 18 18 18 18 18 18 18 18 18 18	(D. 14 e) 12 e 10 e 10 e 10 e 10 e 10 e 10 e 10 e 10	Peak Variation from Setpoint (C): -3.61/2.7 Uniformity (C): 1.85			
14		Stability (C): 2.95			
11 11 10		1-min Door Opening Recovery(min): 6 Cycle on Time(%): 20			
9 1	0	Cycle (on/off) Time (min): 3/12			
8 76	2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2	Energy Consumption (kw-hr/day): 0.491 Heat Rejection Rate (btu/hr): 67			
5	Time (min) Pull Down Time to setpoint (Min) 27				
1 11 21 31 41 51 61 71 81 91 101 111 121 131 141  Pull Doug and Warm Line (min)	At 5C Cycle	Warm Up Time to 15°C (min): 136			
Pull Down and Warm Up  Performance is nominal and individual units may vary.		Sound (dBA): 48.2			

<sup>1)</sup>Performance is nominal and individual units may vary.
2)Product performance will differ due to product amount, product size and operating conditions.
3)Continuous product enhancements may, without notice, result in amendments or ommisions to this specification. Thermo Scientific cannot accept responsibility for damage, injury, loss or expenses resulting from misapplication of the information herein.

<sup>© 2021</sup> Thermo Scientific Inc. All trademarks are the property of Thermo Scientific Inc. and its subsidiaries. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.



## Typical Cabinet Temperature Map Refrigerator 22cuft , 4 Inner-Shelves + Base, Single Outer Door

Temperatures are averages during > 20 cycles after reaching a setpoint of 5C

2: 6.0C

## **Top View of Shelves**

REAR

FRONT

M3 M4 M5 M1 M2

M10

M13 M14

M15

M12

M23 M24 M25 M21 M22


Cabinet Average: 5.1 C
Probe Average: 6.8 C

M11

Peak Variation: +2.3 C / -3.7 C

Front View







20297-A-8-5

	M1	M2	M3	M4	M5	M10	M11	M12	M13
Avg	6.2	6	5.5	5.5	5.9	5.2	5.4	5.2	4.7
Max	7.4	7.4	7.1	7.1	7.3	6.5	6.4	6.4	6
Min	4.3	3.4	2.2	1.6	3.2	3.1	3.6	3.4	2.3

	M14	M15	M20	M21	M22	M23	M24	M25
Avg	4.4	4.9	4.7	5.3	5.1	4.7	4.7	4.7
Max	6	6	5.9	6.3	6.3	5.9	5.8	5.8
Min	1.3	3.1	2.9	3.4	2.6	2.3	2.8	2.5

## **thermo**scientific





**SECTION A-A** 

